

Appendix F

The QA/QC Procedures and Results

HIGH VOLUME AIR SAMPLER
SITE VISIT LOG SHEET

Site Name: R.E. Site No.: AM1
 Date of visit: 12-1-05 Hour of Visit: 13:47
 Staff name: W.L. YAK, H.K. Tsang HVAS S/N: 2198
 Used filter paper no.: LS62 New filter paper no.: LS64
 Type of filter: Glass-fibre

I. Ambient Conditions

Temperature, $T_a = \frac{35.1 + 273}{-308.1}$ K Pressure, $P_a = 1008$ mb

II. Correction of manometer reading

Calibration orifice No.	Manometer reading at site conditions corresponds to $Q_{STD} = 40 \text{ ft}^3/\text{min}$. (inch H ₂ O)
1534(09/2004)	$\Delta H_a = 18.33(T_a/P_a) = 5.62$

Manometer reading before calibration: 5.8
 Adjustment of flow controller (Y/N): Y
 Manometer reading after calibration: 5.6

Note: Tolerance Limit of HVAS flow: $\pm 1.0 \text{ ft}^3/\text{min}$. Corresponding limits for manometer : $\pm 0.2 \text{ inch H}_2\text{O}$

III. General Conditions of HVAS

IV. Remarks

HIGH VOLUME AIR SAMPLER
SITE VISIT LOG SHEET

Site Name: E.G. Site No.: AM2
 Date of visit: 12-P-05 Hour of Visit: 11:45
 Staff name: W.L. MAK, H.K. TSANG HVAS S/N: 2195
C. H. HUNG New filter paper no.: LS65
 Used filter paper no.: LS63
 Type of filter: Glass-fibre

I. Ambient Conditions

Temperature, $T_a = \frac{32.5 + 273}{-305.5}$ K Pressure, $P_a = \frac{1010}{1018}$ mb

II. Correction of manometer reading

Calibration orifice No.	Manometer reading at site conditions corresponds to $Q_{STD} = 40 \text{ ft}^3/\text{min}$. (inch H_2O)
1534(09/2004)	$\Delta H_a = 18.33(T_a/P_a) = \underline{5.54}$

Manometer reading before calibration: 5.8
 Adjustment of flow controller (Y/N): Y
 Manometer reading after calibration: 5.5

Note: Tolerance Limit of HVAS flow: $\pm 1.0 \text{ ft}^3/\text{min}$. Corresponding limits for manometer : $\pm 0.2 \text{ inch H}_2\text{O}$

III. General Conditions of HVAS

IV. Remarks

PARTISOL TSP SAMPLER
SITE VISIT LOG SHEET

Site Name: Ash Lagoon Site Number: Am3
Date of Visit: 12-9-05 Hour of Visit: 14:10
Staff Name: W. L. MAX Partisol S/N: 2000B 20755c410
Used Filter No.: PC76 New Filter No.: PC77
Ambient temperature: 34.8° Ambient pressure: 1009.4/1007

I. General Services

1. Replace control unit Large In-line Filter X
2. Clean the sample inlet head ✓
3. Clean sample tube ✓
4. Clean / Replace pump head X
5. Clean / Replace piston X

II. Operational Audits (3 months interval as recommended by manufacturer)

1. Temperature Check (Ambient temperature $\pm 2^{\circ}\text{C}$)
34.4 °C Before Calibration: Y/N After _____ °C
2. Pressure Check (Ambient pressure ± 20 mbar)(factor = 0.000987)
1008 mbar Before Calibration: Y/N After _____ mbar
3. Flow Check (16.7 \pm 1.1 litre/min)
16.7 l/min Before Calibration: Y/N After _____ l/min

III. Remarks

MINI VOLUME AIR SAMPLER
SITE VISIT LOG SHEET

Site Name: TYV Site No.: AM4
Date of visit: 12-9-05 Hour of Visit: 11:15
Staff name: H.K.TSANG MINIVOL S/N: 383
Used filter paper no.: MH62 New filter paper no.: MH63

Type of filter: ~~Cellulose~~ / Glass-fibre
(Delete as appropriate)

- I. Calibration is performed by using Drycal DC-2 Flow Calibrator
5 Sl/min set point is recommended

5.0 Before 5.0 After

II. General Service of Mini Vol Air Sampler

1. Clean Rotameter: X
2. Clean / replace Pump Valves: X
3. Clean / replace Pump Diaphragms: X
4. Clean Impaction Inlet: ✓
5. Replace Timer Battery Every 6 months: X
6. Replace Inlet Filter: ✓

III. Remarks

THE HONGKONG ELECTRIC CO., LTD.
LAMMA POWER STATION EXTENSION
NOISE MONITORING STATION
SITE VISIT LOG SHEET

Location Ash Lagoon/~~Ching Lam~~*

Date 14-9-05 Time 10:30

Equipment Rion NA-27/~~B&K 2238F~~* Sound Level Meter

Serial Number ~~00111465/00111466/00111467/2343838/2356907*~~

Staff Attended W.L.MAK ; H.K.TSANG

1. Calibration

Acoustic calibrator used Rion NC-74

Calibration level before adjustment (dB(A)) 94.0

Calibration level after adjustment (dB(A)) 94

2. Weather Conditions

a. ~~Sunny/fine/cloudy/showery/heavy rain*~~

b. ~~Strong wind/breeze/calm*~~

3. Remark/Observation

Note: * - Please delete where inappropriate

Equipment Calibration Record for September 2005

Site: Civil works for 275kV Cable Route from Lamma Island to Cyberport

Noise Equipment Used: RION NL-31(*) / ACO-TYPE 6224 (#)

Calibrator Used: ACO-TYPE 2126

Measurement Location: N4 – Pak Kok Tsui No. 2

Date	Calibration Level before Measurement (dB(A))	Calibration Level after Measurement (dB(A))	Calibrated by
02/09/2005 (*)	94.0	94.0	Anthony Wong
06/09/2005 (#)	94.0	94.0	Anthony Wong
09/09/2005 (#)	94.0	94.0	Anthony Wong
13/09/2005 (#)	94.0	94.0	Anthony Wong
16/09/2005 (#)	94.0	94.0	Anthony Wong
20/09/2005 (#)	94.0	94.0	Anthony Wong
23/09/2005 (#)	94.0	94.0	Anthony Wong
27/09/2005 (#)	94.0	94.0	Anthony Wong
30/09/2005 (#)	94.0	94.0	Anthony Wong

Measurement Location: N5 – Pak Kok Tsui No. 8

Date	Calibration Level before Measurement (dB(A))	Calibration Level after Measurement (dB(A))	Calibrated by
02/09/2005 (*)	94.0	94.0	Anthony Wong
06/09/2005 (#)	94.0	94.0	Anthony Wong
09/09/2005 (#)	94.0	94.0	Anthony Wong
13/09/2005 (#)	94.0	94.0	Anthony Wong
16/09/2005 (#)	94.0	94.0	Anthony Wong
20/09/2005 (#)	94.0	94.0	Anthony Wong
23/09/2005 (#)	94.0	94.0	Anthony Wong
27/09/2005 (#)	94.0	94.0	Anthony Wong
30/09/2005 (#)	94.0	94.0	Anthony Wong

Note: Measurement accepted as valid only if the calibration levels from before and after the noise measurement agreed to within 1.0 dB.