

Appendix F

The QA/QC Procedures and Results

HIGH VOLUME AIR SAMPLER
SITE VISIT LOG SHEET

Site Name: R.E. Site No.: AM1
 Date of visit: 12-10-05 Hour of Visit: 11:40
 Staff name: MAK, Fai. HVAS S/N: 2198
 Used filter paper no.: LS 72 New filter paper no.: LS 74
 Type of filter: Glass-fibre

I. Ambient Conditions

Temperature, $T_a = \frac{30.3 + 273}{303.3}$ Pressure, $P_a = 1009$ mb

II. Correction of manometer reading

Calibration orifice No.	Manometer reading at site conditions corresponds to $Q_{STD} = 40$ ft ³ /min. (inch H ₂ O)
1535(09/2005)	$\Delta H_a = 19.29(T_a/P_a) = 5.80$

Manometer reading before calibration: 6.1
 Adjustment of flow controller (Y/N): Y
 Manometer reading after calibration: 5.8

Note: Tolerance Limit of HVAS flow: ± 1.0 ft³/min. Corresponding limits for manometer : ± 0.2 inch H₂O

III. General Conditions of HVAS

IV. Remarks

HIGH VOLUME AIR SAMPLER
SITE VISIT LOG SHEET

Site Name: E.G. Site No.: AM2
 Date of visit: 12/10/05 Hour of Visit: 11:40
 Staff name: MAK, Fai HVAS S/N: 2195
 Used filter paper no.: LS 73 New filter paper no.: LS 75
 Type of filter: Glass-fibre

I. Ambient Conditions

Temperature, $T_a = \frac{30.6 + 273}{303.6}$ K Pressure, $P_a = 1014$ mb

II. Correction of manometer reading

Calibration orifice No.	Manometer reading at site conditions corresponds to $Q_{STD} = 40$ ft ³ /min. (inch H ₂ O)
1535(09/2005)	$\Delta H_a = 19.29(T_a/P_a) = 5.78$

Manometer reading before calibration: 5.8
 Adjustment of flow controller (Y/N): N
 Manometer reading after calibration: 5.8

Note: Tolerance Limit of HVAS flow: ± 1.0 ft³/min. Corresponding limits for manometer : ± 0.2 inch H₂O

III. General Conditions of HVAS

IV. Remarks

PARTISOL TSP SAMPLER
SITE VISIT LOG SHEET

Site Name: ASH LAGOON Site Number: AM 3
Date of Visit: 12-10-2005 Hour of Visit: 1020
Staff Name: W. L. MAK/HK TSANG Partisol S/N: 2000B20755C410
Used Filter No.: PC 81 New Filter No.: PC 82
Ambient temperature: 30.0 Ambient pressure: 1012

I. General Services

1. Replace control unit Large In-line Filter X
2. Clean the sample inlet head ✓
3. Clean sample tube ✓
4. Clean / Replace pump head X
5. Clean / Replace piston X

II. Operational Audits (3 months interval as recommended by manufacturer)

1. Temperature Check (Ambient temperature $\pm 2^{\circ}\text{C}$)

30.5 $^{\circ}\text{C}$ Before Calibration: Y/N $^{\circ}\text{C}$ After

2. Pressure Check (Ambient pressure ± 20 mbar)(factor = 0.000987)

0.997 mbar Before Calibration: Y/N mbar After

3. Flow Check (16.7 \pm 1.1 litre/min)

 l/min Before Calibration: Y/N l/min After

III. Remarks

MINI VOLUME AIR SAMPLER
SITE VISIT LOG SHEET

Site Name: TYU Site No.: AM4
Date of visit: 12-10-2005 Hour of Visit: 09:30
Staff name: K. F Chan MINIVOL S/N: 3393
Used filter paper no.: MH67 New filter paper no.: MH68
Type of filter: Cellulose Glass-fibre
(Delete as appropriate)

I. Calibration is performed by using Drycal DC-2 Flow Calibrator

5 Sl/min set point is recommended

5-081 Before 5-020 After

II. General Service of Mini Vol Air Sampler

1. Clean Rotameter: _____ X
2. Clean / replace Pump Valves: _____ X
3. Clean / replace Pump Diaphragms: _____ X
4. Clean Impaction Inlet: _____ J
5. Replace Timer Battery Every 6 months: _____ X
6. Replace Inlet Filter: _____ J

III. Remarks

THE HONGKONG ELECTRIC CO., LTD.
LAMMA POWER STATION EXTENSION
TEOM 1400A CONTINUOUS DUST MONITOR
DATA QUALITY ASSURANCE LOG SHEET

Month : October Year : 2005

Reservoir (AM1)					
Date	Frequency (Hz) (230 – 260)	Noise (< 0.1)	Operation Mode (Mode 4)	Main Flow (l/min) (0.94 – 1.06)	Aux. Flow (l/min) (14.67 – 16.67)
4/10/2005	256.23	0.041	4	1.00	15.68
10/10/2005	256.07	0.047	4	1.00	15.68
16/10/2005	255.85	0.055	4	1.00	15.68
22/10/2005	255.35	0.030	4	1.00	15.68
28/10/2005	254.66	0.036	4	1.00	15.68

East Gate (AM2)					
Date	Frequency (Hz) (230 – 250)	Noise (< 0.1)	Operation Mode (Mode 4)	Main Flow (l/min) (0.94 – 1.06)	Aux. Flow (l/min) (14.67 – 16.67)
4/10/2005	245.78	0.031	4	0.99	15.63
10/10/2005	245.30	0.063	4	0.99	15.64
16/10/2005	245.14	0.041	4	0.99	15.64
22/10/2005	244.65	0.034	4	1.00	15.63
28/10/2005	245.14	0.039	4	0.99	15.63

Ash Lagoon (AM3)					
Date	Frequency (Hz) (240 – 270)	Noise (< 0.1)	Operation Mode (Mode 4)	Main Flow (l/min) (0.94 – 1.06)	Aux. Flow (l/min) (14.67 – 16.67)
4/10/2005	256.51	0.042	4	1.00	15.69
10/10/2005	256.39	0.040	4	1.00	15.70
16/10/2005	256.17	0.046	4	1.00	15.70
22/10/2005	255.58	0.047	4	1.00	15.69
28/10/2005	254.94	0.030	4	1.00	15.69

Maintenance Record			
	Reservoir	East Gate	Ash Lagoon
TEOM Filter Exchange	✓	✓	✓
Clean TSP Inlet	✓	✓	✓
Replace flow in-line filter			
Pump Repair			
Leak Check			
Flow Audit			
Flow Controller Calibration			
A/C filter cleaning	✓	✓	✓

Remarks:

Prepared by : Alex.

Checked by : [Signature]

THE HONGKONG ELECTRIC CO., LTD.
LAMMA POWER STATION EXTENSION
NOISE MONITORING STATION
SITE VISIT LOG SHEET

Location Ash Lagoon/~~Ching Lam~~*

Date 17-10-2005 Time 11:00

Equipment Rion NA-27/~~B&K 2238F~~* Sound Level Meter

Serial Number 00111465/00111466/00111467/2343838/2356907*

Staff Attended W. L. MAK

1. Calibration

Acoustic calibrator used Rion NC-74

Calibration level before adjustment (dB(A)) 94.0

Calibration level after adjustment (dB(A)) 94

2. Weather Conditions

a. Sunny/~~fine~~/~~cloudy~~/~~showery~~/~~heavy rain~~*

b. ~~Strong wind~~/~~breeze~~/~~calm~~*

3. Remark/Observation

Note: * - Please delete where inappropriate

THE HONGKONG ELECTRIC CO., LTD.
LAMMA POWER STATION EXTENSION
NOISE MONITORING STATION
SITE VISIT LOG SHEET

Location Ash Lagoon/Ching Lam*

Date 20-10-05 Time 16:10

Equipment Rion NA-27/B&K 2238F* Sound Level Meter

Serial Number 00111465/00111466/00111467/2343838/2356907*

Staff Attended W. L. MAK, H. K. TSANG

1. Calibration

Acoustic calibrator used Rion NC-74

Calibration level before adjustment (dB(A)) 94.0

Calibration level after adjustment (dB(A)) 94

2. Weather Conditions

a. ~~Sunny/fine/cloudy/showery/heavy rain*~~

b. ~~Strong wind/breeze/calm*~~

3. Remark/Observation

Note: * - Please delete where inappropriate

18/5/2004

Equipment Calibration Record for October 2005

Site: Civil works for 275kV Cable Route from Lamma Island to Cyberport

Noise Equipment Used: ACO TYPE 6224 (*) / RION NL – 31 (#)

Calibrator Used: ACO TYPE 2126 (*) / RION NC – 74 (#)

Measurement Location: N4 – Pak Kok Tsui No. 2

Date	Calibration Level before Measurement (dB(A))	Calibration Level after Measurement (dB(A))	Calibrated by
04/10/2005(*)	94.0	94.0	Anthony Wong
07/10/2005(*)	94.0	94.0	Anthony Wong
10/10/2005(*)	94.0	94.0	Anthony Wong
14/10/2005(*)	94.0	94.0	Anthony Wong
18/10/2005(*)	94.0	94.0	Anthony Wong
21/10/2005(*)	94.0	94.0	Anthony Wong
25/10/2005(#)	94.0	94.0	Anthony Wong
28/10/2005(#)	94.0	94.0	Anthony Wong

Measurement Location: N5 – Pak Kok Tsui No. 8

Date	Calibration Level before Measurement (dB(A))	Calibration Level after Measurement (dB(A))	Calibrated by
04/10/2005(*)	94.0	94.0	Anthony Wong
07/10/2005(*)	94.0	94.0	Anthony Wong
10/10/2005(*)	94.0	94.0	Anthony Wong
14/10/2005(*)	94.0	94.0	Anthony Wong
18/10/2005(*)	94.0	94.0	Anthony Wong
21/10/2005(*)	94.0	94.0	Anthony Wong
25/10/2005(#)	94.0	94.0	Anthony Wong
28/10/2005(#)	94.0	94.0	Anthony Wong

Note: Measurement accepted as valid only if the calibration levels from before and after the noise measurement agreed to within 1.0 dB.