

Appendix F

The QA/QC Procedures and Results

HIGH VOLUME AIR SAMPLER
SITE VISIT LOG SHEET

Site Name: RTE Site No.: AM1
 Date of visit: 11-3-06 Hour of Visit: 10:00
 Staff name: H.K. BANG HVAS S/N: 2198
 Used filter paper no.: LT 27 New filter paper no.: LT 29
 Type of filter: Glass-fibre

I. Ambient Conditions

Temperature, $T_a = \overset{22.6+273}{295.6}$ K Pressure, $P_a = \underline{1010}$ mb

II. Correction of manometer reading

Calibration orifice No.	Manometer reading at site conditions corresponds to $Q_{STD} = 40 \text{ ft}^3/\text{min}$. (inch H ₂ O)
1535(09/2005)	$\Delta H_a = 19.29(T_a/P_a) = \underline{5.65}$

Manometer reading before calibration: 6.0
 Adjustment of flow controller (Y/N): Y
 Manometer reading after calibration: 5.7

Note: Tolerance Limit of HVAS flow: $\pm 1.0 \text{ ft}^3/\text{min}$. Corresponding limits for manometer : $\pm 0.2 \text{ inch H}_2\text{O}$

III. General Conditions of HVAS

IV. Remarks

HIGH VOLUME AIR SAMPLER
SITE VISIT LOG SHEET

Site Name: E.G Site No.: AM2
 Date of visit: 16-3-06 Hour of Visit: 15:10
 Staff name: H.K.TSANG HVAS S/N: 2195
 Used filter paper no.: LT28 New filter paper no.: LT30
 Type of filter: Glass-fibre

I. Ambient Conditions

Temperature, $T_a = \frac{24+273}{273}$ K Pressure, $P_a = 1010$ mb

II. Correction of manometer reading

Calibration orifice No.	Manometer reading at site conditions corresponds to $Q_{STD} = 40 \text{ ft}^3/\text{min}$. (inch H_2O)
1535(09/2005)	$\Delta H_a = 19.29(T_a/P_a) = 5.67$

Manometer reading before calibration: 5.6
 Adjustment of flow controller (Y/N): N
 Manometer reading after calibration: 5.6

Note: Tolerance Limit of HVAS flow: $\pm 1.0 \text{ ft}^3/\text{min}$. Corresponding limits for manometer : $\pm 0.2 \text{ inch H}_2\text{O}$

III. General Conditions of HVAS

IV. Remarks

PARTISOL TSP SAMPLER
SITE VISIT LOG SHEET

Site Name: Ash Lagoon Site Number: AM3
Date of Visit: 10-3-06 Hour of Visit: 10:30
Staff Name: W.H. MAK, H.K. ISANH Partisol S/N: 2000B207330410
Used Filter No.: PD10 New Filter No.: PD11
Ambient temperature: 20°C Ambient pressure: 1010 mbar

I. General Services

1. Replace control unit Large In-line Filter x
2. Clean the sample inlet head x
3. Clean sample tube ✓
4. Clean / Replace pump head ✓
5. Clean / Replace piston x

II. Operational Audits (3 months interval as recommended by manufacturer)

1. Temperature Check (Ambient temperature $\pm 2^\circ\text{C}$)

21 °C Before Calibration: Y/N 21 °C After

2. Pressure Check (Ambient pressure ± 20 mbar)(factor = 0.000987)

1016 mbar Before Calibration: Y/N 1016 mbar After

3. Flow Check (16.7 \pm 1.1 litre/min)

16.8 l/min Before Calibration: Y/N 16.8 l/min After

III. Remarks

MINI VOLUME AIR SAMPLER

SITE VISIT LOG SHEET

Site Name: TYV Site No.: AM4

Date of visit: 16-3-06 Hour of Visit: 16:10

Staff name: H. K. TSANG MINTVOL S/N: 3393

Used filter paper no.: MH95 New filter paper no.: MH96

Type of filter: ~~Cellulose~~ / Glass-fibre
(Delete as appropriate)

I. Calibration is performed by using Drycal DC-2 Flow Calibrator

5 SI/min set point is recommended

5.0 Before 5.0 After

II. General Service of Mini Vol Air Sampler

1. Clean Rotameter: ✓
2. Clean / replace Pump Valves: ✓
3. Clean / replace Pump Diaphragms: x
4. Clean Impaction Inlet: ✓
5. Replace Timer Battery Every 6 months: x
6. Replace Inlet Filter: ✓

III. Remarks

**THE HONGKONG ELECTRIC CO., LTD.
LAMMA POWER STATION EXTENSION
TEOM 1400A CONTINUOUS DUST MONITOR
DATA QUALITY ASSURANCE LOG SHEET**

Month : March Year : 2006

Reservoir (AM1)					
Date	Frequency (Hz) (230 – 260)	Noise (< 0.1)	Operation Mode (Mode 4)	Main Flow (l/min) (0.94 – 1.06)	Aux. Flow (l/min) (14.67 – 16.67)
3/3/2006	257.15	0.032	4	1.00	15.68
9/3/2006	256.86	0.050	4	1.00	15.68
15/3/2006	256.61	0.044	4	1.00	15.68
21/3/2006	256.01	0.021	4	1.00	15.68
27/3/2006	255.77	0.057	4	1.00	15.68

East Gate (AM2)					
Date	Frequency (Hz) (230 – 250)	Noise (< 0.1)	Operation Mode (Mode 4)	Main Flow (l/min) (0.94 – 1.06)	Aux. Flow (l/min) (14.67 – 16.67)
3/3/2006	245.52	0.054	4	0.99	15.64
9/3/2006	245.23	0.042	4	0.99	15.62
15/3/2006	244.99	0.023	4	0.99	15.62
21/3/2006	244.40	0.055	4	0.99	15.64
27/3/2006	244.15	0.043	4	0.99	15.63

Ash Lagoon (AM3)					
Date	Frequency (Hz) (240 – 270)	Noise (< 0.1)	Operation Mode (Mode 4)	Main Flow (l/min) (0.94 – 1.06)	Aux. Flow (l/min) (14.67 – 16.67)
3/3/2006	247.68	0.061	4	0.99	15.67
9/3/2006	247.43	0.042	4	1.00	15.20
15/3/2006	248.01	0.030	4	1.00	15.66
21/3/2006	247.64	0.037	4	1.00	15.67
27/3/2006	247.42	0.055	4	1.00	15.68

Maintenance Record			
	Reservoir	East Gate	Ash Lagoon
TEOM Filter Exchange	✓	✓	✓
Clean TSP Inlet	✓	✓	✓
Replace flow in-line filter			
Pump Repair			
Leak Check			
Flow Audit	✓	✓	✓
Flow Controller Calibration			
A/C filter cleaning	✓	✓	✓

Remarks:

Prepared by :

Checked by :

THE HONGKONG ELECTRIC CO., LTD.
LAMMA POWER STATION EXTENSION
NOISE MONITORING STATION
SITE VISIT LOG SHEET

Location Ash Lagoon/~~Ching Lam~~*

Date 14-3-06 Time 16:00

Equipment Rion NA-27/~~B&K-2238F~~* Sound Level Meter

Serial Number ~~00111465/00111466/00111467/2343838/2356907*~~

Staff Attended U.L. MAK ; H.K. TSANG

1. Calibration

Acoustic calibrator used Rion NC-74

Calibration level before adjustment (dB(A)) 94.0

Calibration level after adjustment (dB(A)) 94

2. Weather Conditions

a. ~~Sunny/fine/cloudy/showery/heavy rain*~~

b. ~~Strong wind/breeze/calm*~~

3. Remark/Observation

Note: * - Please delete where inappropriate

THE HONGKONG ELECTRIC CO., LTD.
LAMMA POWER STATION EXTENSION
NOISE MONITORING STATION
SITE VISIT LOG SHEET

Location Ash Lagoon/Ching Lam*

Date 20-3-06 Time 10:30

Equipment Rion NA-27/B&K 2238F* Sound Level Meter

Serial Number 00111465/00111466/00111467/2343838/2356907*

Staff Attended W.L.MAK ; H.K.TSANG

1. Calibration

Acoustic calibrator used Rion NC-74

Calibration level before adjustment (dB(A)) 94.0

Calibration level after adjustment (dB(A)) 94

2. Weather Conditions

a. ~~Sunny/fine/cloudy/showery/heavy rain*~~

b. ~~Strong wind/breeze/calm*~~

3. Remark/Observation

Note: * - Please delete where inappropriate

Equipment Calibration Record for March 2006

Site: Civil works for 275kV Cable Route from Lamma Island to Cyberport

Noise Equipment Used: RION NL – 31

Calibrator Used: RION NC – 74

Measurement Location: N4 – Pak Kok Tsui No. 2

Date	Calibration Level before Measurement (dB(A))	Calibration Level after Measurement (dB(A))	Calibrated by
03/03/2006	94.0	94.0	C K Law
07/03/2006	94.0	94.0	C K Law
10/03/2006	94.0	94.0	C K Law
14/03/2006	94.0	94.0	C K Law
17/03/2006	94.0	94.0	C K Law
21/03/2006	94.0	94.0	C K Law
24/03/2006	94.0	94.0	C K Law
28/03/2006	94.0	94.0	C K Law
31/03/2006	94.0	94.0	C K Law

Measurement Location: N5 – Pak Kok Tsui No. 8

Date	Calibration Level before Measurement (dB(A))	Calibration Level after Measurement (dB(A))	Calibrated by
03/03/2006	94.0	94.0	C K Law
07/03/2006	94.0	94.0	C K Law
10/03/2006	94.0	94.0	C K Law
14/03/2006	94.0	94.0	C K Law
17/03/2006	94.0	94.0	C K Law
21/03/2006	94.0	94.0	C K Law
24/03/2006	94.0	94.0	C K Law
28/03/2006	94.0	94.0	C K Law
31/03/2006	94.0	94.0	C K Law

Note: Measurement accepted as valid only if the calibration levels from before and after the noise measurement agreed to within 1.0 dB.