

Appendix E

The QA/QC Procedures and Results

Equipment Calibration Record for December 2006

Site: Civil works for 275kV Cable Route from Lamma Island to Cyberport

Noise Equipment Used: Rion NL-31

Calibrator Used: Rion NC-74

Measurement Location: N4 – Pak Kok Tsui No. 2

Date	Calibration Level before Measurement (dB(A))	Calibration Level after Measurement (dB(A))	Calibrated by
01/12/2006	94.0	94.0	C K Siu
05/12/2006	94.0	94.0	Anthony Tang
08/12/2006	94.0	94.0	Anthony Tang
12/12/2006	94.0	94.0	C K Siu
15/12/2006	94.0	94.0	C K Siu
19/12/2006	94.0	94.0	C K Siu
22/12/2006	94.0	94.0	C K Siu
27/12/2006	94.0	94.0	Anthony Tang
29/12/2006	94.0	94.0	Anthony Tang

Measurement Location: N5 – Pak Kok Tsui No. 8

Date	Calibration Level before Measurement (dB(A))	Calibration Level after Measurement (dB(A))	Calibrated by
01/12/2006	94.0	94.0	C K Siu
05/12/2006	94.0	94.0	Anthony Tang
08/12/2006	94.0	94.0	Anthony Tang
12/12/2006	94.0	94.0	C K Siu
15/12/2006	94.0	94.0	C K Siu
19/12/2006	94.0	94.0	C K Siu
22/12/2006	94.0	94.0	C K Siu
27/12/2006	94.0	94.0	Anthony Tang
29/12/2006	94.0	94.0	Anthony Tang

Note: Measurement accepted as valid only if the calibration levels from before and after the noise measurement agreed to within 1.0 dB.